



Jeeter
2626 South Roosevelt Street
Tempe, AZ 85282
License #: 00000066DCBO00410690
Sample ID: 2508SMAZ1450.4260
Batch #: DFAZ-WATZKZ-082725



SMITHERS

CERTIFICATE OF ANALYSIS
License #: 00000020LCVT89602592

Certificate: 15991

(Jeeter) INFUSED WATERMELON ZKZ PREROLL 0.5G X 5, 1G, 2G

Batch #: DFAZ-WATZKZ-082725

Strain: Watermelon Zkz

Parent Batch #:

Production Method: Other

Harvest Date: 04/28/2025

Received: 08/28/2025

Sample ID: 2508SMAZ1450.4260

Amount Received: 12.2 g

Sample Type: Enhanced/Infused Flowers

Sample Collected: 08/28/2025 12:31:00

Manufacture Date: 08/27/2025

Published: 09/03/2025



COMPLIANCE FOR RETAIL

Regulated Analytes

Cannabinoid Profile (Q3)

Tested

Microbial Contaminants

Pass

Residual Solvents

Pass

39.9046%

Total THC

Pesticides, Fungicides,
and Growth Regulators

Pass

Mycotoxins

Pass

Heavy Metals

Pass

ND

Total CBD

Additional Analytes (Not Regulated)

Terpenes Total (Q3)

Tested

Moisture Analysis (Q3)

Tested

Water Activity (Q3)

Tested

ND

CBN

Filth & Foreign (Q3)

Not Tested

Homogeneity (Q3)

Not Tested

Additional Microbial
Contaminants (Q3)

Not Tested

ND

CBG

45.5595%


Total Cannabinoids (Q3)

Ahmed Munshi

Technical Laboratory Director

Smithers CTS Arizona LLC

734 W Highland Avenue, 2nd Floor
Phoenix, AZ 85013
(602) 806-6930



Accreditation #: 103104

The product associated with this COA has been tested by Smithers CTS Arizona LLC, using validated state certified testing methodologies as required by Arizona state law.

Testing results were obtained according to Smithers' quality assurance plan and requirements found in R9-17-404.03 and R9-17-404.04.

This COA is governed by the terms and conditions listed on: <https://www.smithers.com/arizona-terms-conditions>



Jeeter
2626 South Roosevelt Street
Tempe, AZ 85282
License #: 00000066DCBO00410690
Sample ID: 2508SMAZ1450.4260
Batch #: DFAZ-WATZKZ-082725



SMITHERS

CERTIFICATE OF ANALYSIS
License #: 00000020LCVT89602592

Certificate: 15991

Cannabinoid Profile

HPLC

Tested

Sample Prep

Batch Date: 08/29/2025
SOP: 418.AZ
Batch Number: 3992
Test ID: 87826

Sample Analysis

Date: 09/02/2025
SOP: 417.AZ - HPLC
Sample Weight: 0.0408 g
Volume: 40 mL

Analyte	LOD (mg/g)	LOQ (mg/g)	Dil.	Actual % (w/w)	mg/g	Qualifier
CBC	0.3160	0.9580	1	<LOQ	<LOQ	
CBD	0.3160	0.9580	1	ND	ND	
CBDA	0.3160	0.9580	1	ND	ND	
CBDV	0.3160	0.9580	1	ND	ND	
CBG	0.3160	0.9580	1	ND	ND	
CBGA	0.3160	0.9580	1	0.5275	5.2750	
CBN	0.3160	0.9580	1	ND	ND	
d8-THC	0.3160	0.9580	1	ND	ND	
d9-THC	0.3160	0.9580	1	3.3463	33.4630	
THCA	0.3160	0.9580	1	41.6857	416.8570	
THCV	0.3160	0.9580	1	<LOQ	<LOQ	


Cannabinoid Totals	Actual % (w/w)	mg/g	Qualifier
Total THC	39.9046	399.0460	
Total CBD	ND	ND	
Total Cannabinoids	45.5595	455.5950	Q3

Total THC = THC + (0.877 x THCA) and Total CBD = CBD + (0.877 x CBDA)
ND = Not Detected, NT = Not Tested, <LOQ = Below Limit of Quantitation

Ahmed Munshi

Technical Laboratory Director

Smithers CTS Arizona LLC
734 W Highland Avenue, 2nd Floor
Phoenix, AZ 85013
(602) 806-6930



Accreditation #: 103104

The product associated with this COA has been tested by Smithers CTS Arizona LLC, using validated state certified testing methodologies as required by Arizona state law. Testing results were obtained according to Smithers' quality assurance plan and requirements found in R9-17-404.03 and R9-17-404.04. This COA is governed by the terms and conditions listed on: <https://www.smithers.com/arizona-terms-conditions>



Jeeter
2626 South Roosevelt Street
Tempe, AZ 85282
License #: 00000066DCBO00410690
Sample ID: 2508SMAZ1450.4260
Batch #: DFAZ-WATZKZ-082725



SMITHERS

CERTIFICATE OF ANALYSIS
License #: 00000020LCVT89602592

Certificate: 15991

Terpene Total

GC-FID

Tested (4.8890%)


Sample Prep

Batch Date: 09/02/2025
SOP: 419
Batch Number: 4002

Sample Analysis

Date: 09/03/2025
SOP: 419 - GC-FID
Sample Weight: 0.428 g
Volume: 10 mL

Analyte	LOD / LOQ (%)	Dil.	Results (%)	Qualifier	Analyte	LOD / LOQ (%)	Dil.	Results (%)	Qualifier
alpha-Bisabolol	0.0009 / 0.0028	1	0.2618	Q3	gamma-Terpinene	0.0009 / 0.0028	1	<LOQ	Q3
alpha-Cedrene	0.0009 / 0.0028	1	0.0111	Q3	Geraniol	0.0009 / 0.0028	1	0.0209	Q3
alpha-Humulene	0.0009 / 0.0028	1	0.1916	Q3	Geranyl acetate	0.0009 / 0.0028	1	ND	Q3
alpha-Phellandrene	0.0009 / 0.0028	1	ND	Q3	Guaiol	0.0009 / 0.0028	1	<LOQ	Q3
alpha-Pinene	0.0009 / 0.0028	1	0.0663	Q3	Hexahydrothymol	0.0009 / 0.0028	1	ND	Q3
alpha-Terpinene	0.0009 / 0.0028	1	ND	Q3	Isoborneol	0.0009 / 0.0028	1	0.0040	Q3
beta-Myrcene	0.0009 / 0.0028	1	0.4375	Q3	Isopulegol	0.0009 / 0.0028	1	ND	Q3
beta-Pinene	0.0009 / 0.0028	1	0.1331	Q3	Limonene	0.0009 / 0.0028	1	0.7485	Q3
Borneol	0.0009 / 0.0028	1	0.0058	Q3	Linalool	0.0009 / 0.0028	1	0.5855	Q3
Camphene	0.0009 / 0.0028	1	0.0211	Q3	Nerol	0.0009 / 0.0028	1	0.0386	Q3
Camphor	0.0009 / 0.0028	1	0.0098	Q3	Pulegone (+)	0.0009 / 0.0028	1	ND	Q3
3-Carene	0.0009 / 0.0028	1	ND	Q3	Sabinene Hydrate	0.0009 / 0.0028	1	ND	Q3
Caryophyllene oxide	0.0009 / 0.0028	1	0.3199	Q3	Terpineol	0.0009 / 0.0028	1	0.2810	Q3
Cedrol	0.0009 / 0.0028	1	0.0919	Q3	Terpinolene	0.0009 / 0.0028	1	0.0051	Q3
cis-Nerolidol	0.0009 / 0.0028	1	ND	Q3	trans-Caryophyllene	0.0009 / 0.0028	1	1.4571	Q3
cis-Ocimene	0.0009 / 0.0028	1	ND	Q3	trans-Nerolidol	0.0009 / 0.0028	1	ND	Q3
Fenchyl alcohol	0.0009 / 0.0028	1	0.2370	Q3	trans-Ocimene	0.0009 / 0.0028	1	ND	Q3
Eucalyptol	0.0009 / 0.0028	1	ND	Q3	Valencene	0.0009 / 0.0028	1	ND	Q3
Fenchone	0.0009 / 0.0028	1	ND	Q3					



Ahmed Munshi

Technical Laboratory Director

Smithers CTS Arizona LLC

734 W Highland Avenue, 2nd Floor
Phoenix, AZ 85013
(602) 806-6930



Accreditation #: 103104

The product associated with this COA has been tested by Smithers CTS Arizona LLC, using validated state certified testing methodologies as required by Arizona state law. Testing results were obtained according to Smithers' quality assurance plan and requirements found in R9-17-404.03 and R9-17-404.04. This COA is governed by the terms and conditions listed on: <https://www.smithers.com/arizona-terms-conditions>



Jeeter
2626 South Roosevelt Street
Tempe, AZ 85282
License #: 00000066DCBO00410690
Sample ID: 2508SMAZ1450.4260
Batch #: DFAZ-WATZKZ-082725



CERTIFICATE OF ANALYSIS

License #: 00000020LCVT89602592

Certificate: 15991

Microbial Analysis

Pass

Sample Prep

Batch Date: 09/02/2025
SOP: 412.AZ
Batch Number: 4005
Test ID: 87845

Sample Analysis

Date: 09/03/2025
SOP: 412.AZ - 3M Petrifilm
Sample Weight: 1.017 g

Analyte	Allowable Criteria	Actual Result	Pass/Fail	Qualifier
E. coli	< 100 CFU/g	< 100 CFU/g	Pass	

Sample Prep

Batch Date: 09/02/2025
SOP: 406.AZ
Batch Number: 4004
Test ID: 87847

Sample Analysis

Date: 09/03/2025
SOP: 406.AZ - qPCR (MG)
Sample Weight: 1.019 g

Analyte	Allowable Criteria	Actual Result	Pass/Fail	Qualifier
Salmonella	Not Detected in One Gram	Not Detected in One Gram	Pass	

Sample Prep

Batch Date: 09/02/2025
SOP: 406.AZ
Batch Number: 4004
Test ID: 87849

Sample Analysis

Date: 09/03/2025
SOP: 406.AZ - qPCR (MG)
Sample Weight: 1.019 g

Analyte	Allowable Criteria	Actual Result	Pass/Fail	Qualifier
Aspergillus flavus	Not Detected in One Gram	Not Detected in One Gram	Pass	
Aspergillus fumigatus	Not Detected in One Gram	Not Detected in One Gram	Pass	
Aspergillus niger	Not Detected in One Gram	Not Detected in One Gram	Pass	
Aspergillus terreus	Not Detected in One Gram	Not Detected in One Gram	Pass	

Ahmed Munshi

Technical Laboratory Director

Smithers CTS Arizona LLC
734 W Highland Avenue, 2nd Floor
Phoenix, AZ 85013
(602) 806-6930



The product associated with this COA has been tested by Smithers CTS Arizona LLC, using validated state certified testing methodologies as required by Arizona state law. Testing results were obtained according to Smithers' quality assurance plan and requirements found in R9-17-404.03 and R9-17-404.04. This COA is governed by the terms and conditions listed on: <https://www.smithers.com/arizona-terms-conditions>



Jeeter
2626 South Roosevelt Street
Tempe, AZ 85282
License #: 00000066DCBO00410690
Sample ID: 2508SMAZ1450.4260
Batch #: DFAZ-WATZKZ-082725



SMITHERS

CERTIFICATE OF ANALYSIS
License #: 00000020LCVT89602592

Certificate: 15991

Residual Solvents

HS-GC-MS

Pass

Sample Prep

Batch Date: 08/29/2025
SOP: 405.AZ
Batch Number: 3989
Test ID: 87827

Sample Analysis

Date: 09/02/2025
SOP: 405.AZ - HS-GC-MS
Sample Weight: 0.0515 g

Analyte	LOD / LOQ (ppm)	Dil.	Action Limit (ppm)	Results (ppm)	Qualifier	Analyte	LOD / LOQ (ppm)	Dil.	Action Limit (ppm)	Results (ppm)	Qualifier
Acetone	64 / 194	1	1000	ND		Heptane	324 / 971	1	5000	ND	
Acetonitrile	27 / 80	1	410	ND		Hexanes	47 / 141	1	290	ND	
Benzene	0.14 / 0.39	1	2	ND		Isopropyl acetate	324 / 971	1	5000	ND	
Butanes	161 / 485	1	5000	ND		Methanol	194 / 583	1	3000	<LOQ	
Chloroform	4 / 12	1	60	ND		Pentanes	324 / 971	1	5000	ND	
Dichloromethane	39 / 117	1	600	ND		2-Propanol (IPA)	324 / 971	1	5000	ND	
Ethanol	324 / 971	1	5000	ND		Toluene	58 / 173	1	890	ND	
Ethyl acetate	324 / 971	1	5000	ND		Xylenes	282 / 843	1	2170	ND	
Ethyl ether	324 / 971	1	5000	ND							

Ahmed Munshi

Technical Laboratory Director

Smithers CTS Arizona LLC

734 W Highland Avenue, 2nd Floor
Phoenix, AZ 85013
(602) 806-6930



Accreditation #: 103104

The product associated with this COA has been tested by Smithers CTS Arizona LLC, using validated state certified testing methodologies as required by Arizona state law. Testing results were obtained according to Smithers' quality assurance plan and requirements found in R9-17-404.03 and R9-17-404.04. This COA is governed by the terms and conditions listed on: <https://www.smithers.com/arizona-terms-conditions>



Jeeter
2626 South Roosevelt Street
Tempe, AZ 85282
License #: 00000066DCBO00410690
Sample ID: 2508SMAZ1450.4260
Batch #: DFAZ-WATZKZ-082725



CERTIFICATE OF ANALYSIS
License #: 00000020LCVT89602592

Certificate: 15991

Heavy Metals

ICP-MS **Pass**

Sample Prep

Batch Date: 09/02/2025
SOP: 428.AZ
Batch Number: 4001
Test ID: 87828

Sample Analysis

Date: 09/02/2025
SOP: 428.AZ - ICP-MS
Sample Weight: 0.227 g
Volume: 6 mL

Analyte	LOD (ppm)	LOQ (ppm)	Dil.	Action Limit (ppm)	Results (ppm)	Qualifier
Arsenic	0.053	0.176	10	0.4	ND	
Cadmium	0.053	0.176	10	0.4	ND	
Lead	0.053	0.441	10	1	ND	
Mercury	0.053	0.088	10	0.2	ND	

Mycotoxin Analysis

LC-MS/MS **Pass**

Sample Prep

Batch Date: 09/02/2025
SOP: 432.AZ
Batch Number: 3996
Test ID: 87832

Sample Analysis

Date: 09/03/2025
SOP: 424.AZ - LC-MS/MS
Sample Weight: 0.550 g
Volume: 12.5 mL

Analyte	LOD (ppb)	LOQ (ppb)	Dil.	Action Limit (ppb)	Results (ppb)	Qualifier
Total Aflatoxins	3.64	9.09	1	20	ND	R1 V1
Aflatoxin B1	3.64	9.09	1		ND	
Aflatoxin B2	3.64	9.09	1		ND	I1, R1 V1
Aflatoxin G1	3.64	9.09	1		ND	I1
Aflatoxin G2	3.64	4.55	1		ND	
Ochratoxin A	9.09	9.09	1	20	ND	I1 V1

Ahmed Munshi

Technical Laboratory Director

Smithers CTS Arizona LLC
734 W Highland Avenue, 2nd Floor
Phoenix, AZ 85013
(602) 806-6930



The product associated with this COA has been tested by Smithers CTS Arizona LLC, using validated state certified testing methodologies as required by Arizona state law. Testing results were obtained according to Smithers' quality assurance plan and requirements found in R9-17-404.03 and R9-17-404.04. This COA is governed by the terms and conditions listed on: <https://www.smithers.com/arizona-terms-conditions>



Jeeter
2626 South Roosevelt Street
Tempe, AZ 85282
License #: 00000066DCBO00410690
Sample ID: 2508SMAZ1450.4260
Batch #: DFAZ-WATZKZ-082725



SMITHERS

CERTIFICATE OF ANALYSIS
License #: 00000020LCVT89602592

Certificate: 15991

**Pesticides, Fungicides, and
Growth Regulators**
LC-MS/MS **Pass**

Sample Prep

Batch Date: 09/02/2025
SOP: 432.AZ
Batch Number: 3996
Test ID: 87831

Sample Analysis

Date: 09/03/2025
SOP: 424.AZ - LC-MS/MS
Sample Weight: 0.550 g
Volume: 12.5 mL

Analyte	LOD / LOQ (ppm)	Dil.	Action Limit (ppm)	Results (ppm)	Qualifier	Analyte	LOD / LOQ (ppm)	Dil.	Action Limit (ppm)	Results (ppm)	Qualifier
Abamectin B1a	0.075 / 0.227	1	0.5	ND		Hexythiazox	0.152 / 0.455	1	1	ND	
Acephate	0.061 / 0.182	1	0.4	ND		Imazalil	0.030 / 0.091	1	0.2	ND	
Acetamiprid	0.030 / 0.091	1	0.2	ND		Imidacloprid	0.061 / 0.182	1	0.4	ND	
Aldicarb	0.061 / 0.182	1	0.4	ND		Kresoxim-methyl	0.061 / 0.182	1	0.4	ND	
Azoxystrobin	0.030 / 0.091	1	0.2	ND		Malathion	0.030 / 0.091	1	0.2	ND	L1
Bifenazate	0.030 / 0.091	1	0.2	ND		Metaxyl	0.030 / 0.091	1	0.2	ND	
Bifenthrin	0.030 / 0.091	1	0.2	ND	V1, V1	Methiocarb	0.030 / 0.091	1	0.2	ND	
Boscalid	0.061 / 0.182	1	0.4	ND		Methomyl	0.061 / 0.182	1	0.4	ND	
Carbaryl	0.030 / 0.091	1	0.2	ND		Myclobutanil	0.030 / 0.091	1	0.2	ND	V1
Carbofuran	0.030 / 0.091	1	0.2	ND		Naled	0.075 / 0.227	1	0.5	ND	
Chlorantraniliprole	0.030 / 0.091	1	0.2	ND		Oxamyl	0.152 / 0.455	1	1	ND	
Chlorfenapyr	0.152 / 0.455	1	1	ND		Paclobutrazol	0.061 / 0.182	1	0.4	ND	V1
Chlorpyrifos	0.030 / 0.091	1	0.2	ND		Permethrins	0.030 / 0.091	1	0.2	ND	
Clofentezine	0.030 / 0.091	1	0.2	ND		Phosmet	0.030 / 0.091	1	0.2	ND	
Cyfluthrin	0.152 / 0.455	1	1	ND		Piperonyl Butoxide	0.303 / 0.909	1	2	ND	
Cypermethrin	0.152 / 0.455	1	1	ND		Prallethrin	0.030 / 0.091	1	0.2	ND	
Daminozide	0.152 / 0.455	1	1	ND		Propiconazole	0.061 / 0.182	1	0.4	ND	
Diazinon	0.030 / 0.091	1	0.2	ND	R1	Propoxur	0.030 / 0.091	1	0.2	ND	
Dichlorvos	0.015 / 0.045	1	0.1	ND		Pyrethrins	0.127 / 0.381	1	1	ND	
Dimethoate	0.030 / 0.091	1	0.2	ND		Pyridaben	0.030 / 0.091	1	0.2	ND	
Ethoprophos	0.030 / 0.091	1	0.2	ND		Spinosad	0.030 / 0.091	1	0.2	ND	
Etofenprox	0.061 / 0.182	1	0.4	ND		Spiromesifen	0.030 / 0.091	1	0.2	ND	
Etoxazole	0.030 / 0.091	1	0.2	ND		Spirotetramat	0.030 / 0.091	1	0.2	ND	
Fenoxycarb	0.030 / 0.091	1	0.2	ND	L1	Spiroxamine	0.061 / 0.182	1	0.4	ND	
Fenpyroximate	0.061 / 0.182	1	0.4	ND		Tebuconazole	0.061 / 0.182	1	0.4	ND	
Fipronil	0.061 / 0.182	1	0.4	ND		Thiacloprid	0.030 / 0.091	1	0.2	ND	
Fonicamid	0.152 / 0.455	1	1	ND	I1	Thiamethoxam	0.030 / 0.091	1	0.2	ND	
Fludioxonil	0.061 / 0.182	1	0.4	ND		Trifloxystrobin	0.030 / 0.091	1	0.2	ND	V1

Ahmed Munshi

Technical Laboratory Director

Ahmed Munshi

Smithers CTS Arizona LLC

734 W Highland Avenue, 2nd Floor
Phoenix, AZ 85013
(602) 806-6930



Accreditation #: 103104

The product associated with this COA has been tested by Smithers CTS Arizona LLC, using validated state certified testing methodologies as required by Arizona state law. Testing results were obtained according to Smithers' quality assurance plan and requirements found in R9-17-404.03 and R9-17-404.04. This COA is governed by the terms and conditions listed on: <https://www.smithers.com/arizona-terms-conditions>



Jeeter
2626 South Roosevelt Street
Tempe, AZ 85282
License #: 00000066DCBO00410690
Sample ID: 2508SMAZ1450.4260
Batch #: DFAZ-WATZKZ-082725



SMITHERS

CERTIFICATE OF ANALYSIS

License #: 00000020LCVT89602592

Certificate: 15991

Moisture Analysis

Moisture: **8.75 %**
Qualifier: Q3

Sample Prep and Analysis

Date: 08/28/2025
SOP: 411 - Moisture Balance
Sample Weight: 0.320 g

Water Activity

Water Activity: **0.4668AW aw**
Qualifier: Q3


Sample Prep and Analysis

Date: 08/28/2025
SOP: 433 - Water Activity Meter
Sample Weight: 0.654 g

Ahmed Munshi

Technical Laboratory Director

Smithers CTS Arizona LLC
734 W Highland Avenue, 2nd Floor
Phoenix, AZ 85013
(602) 806-6930



Accreditation #: 103104

The product associated with this COA has been tested by Smithers CTS Arizona LLC, using validated state certified testing methodologies as required by Arizona state law. Testing results were obtained according to Smithers' quality assurance plan and requirements found in R9-17-404.03 and R9-17-404.04. This COA is governed by the terms and conditions listed on: <https://www.smithers.com/arizona-terms-conditions>



Jeeter
2626 South Roosevelt Street
Tempe, AZ 85282
License #: 00000066DCBO00410690
Sample ID: 2508SMAZ1450.4260
Batch #: DFAZ-WATZKZ-082725



SMITHERS

CERTIFICATE OF ANALYSIS
License #: 00000020LCVT89602592

Certificate: 15991

Qualifier Legend

- B1** The target analyte detected in the calibration is at or above the limit of quantitation, but the sample result for potency testing, is below the limit of quantitation.
- B2** The target analyte detected in the calibration blank, or the method blank is at or above the limit of quantitation, but the sample result when testing for pesticides, fungicides, herbicides, growth regulators, heavy metals, or residual solvents, is below the maximum allowable concentration for the analyte.
- D1** The limit of quantitation and the sample results were adjusted to reflect sample dilution.
- I1** The relative intensity of a characteristic ion in a sample analyte exceeded the acceptance with respect to the reference spectra, indicating interference.
- L1** When testing for pesticides, fungicides, herbicides, growth regulators, heavy metals, or residual solvents, the percent recovery of a laboratory control sample is greater than the acceptance limits, but the sample's target analytes were not detected above the maximum allowable concentrations for the analytes in the sample.
- M1** The recovery from the matrix spike was high, but the recovery from the laboratory control sample was within acceptance criteria.
- M2** The recovery from the matrix spike was low, but the recovery from the laboratory control sample was within acceptance criteria.
- M3** The recovery from the matrix spike was unusable because the analyte concentration was disproportionate to the spike level, but the recovery from the laboratory control sample was within acceptance criteria.
- M4** The analysis of a spiked sample required a dilution such that the spike recovery calculation does not provide useful information, but the recovery from the associated laboratory control sample was within acceptance criteria.
- M5** The analyte concentration was determined by the method of standard addition, in which the standard is added directly to the aliquots of the analyzed sample.
- M6** A description of the variance is described in the final report of testing according to R9-17-404.06(B)(3)(d)(ii).
- Q1** Sample integrity was not maintained.
- Q2** The sample is heterogeneous, and sample homogeneity could not be readily achieved using routine laboratory practices.
- Q3** Testing result is for informational purposes only and cannot be used to satisfy dispensary testing requirements in R9-17-317.01(A) or labeling requirements in R9-17-317.
- R1** The relative percent difference for the laboratory control sample and duplicate exceeded the limit, but the recovery was within acceptance criteria.
- R2** The relative percent difference for a sample and duplicate exceeded the limit.
- V1** The recovery from continuing calibration verification standards exceeded the acceptance limits, but the sample's target analytes were not detected above the maximum allowable for the analytes in the sample.

Cultivated By:

Manufactured By:


Disclaimer: Using marijuana during pregnancy could cause birth defects or other health issues to your unborn child.

Ahmed Munshi

Technical Laboratory Director

Smithers CTS Arizona LLC

734 W Highland Avenue, 2nd Floor
Phoenix, AZ 85013
(602) 806-6930



Accreditation #: 103104

The product associated with this COA has been tested by Smithers CTS Arizona LLC, using validated state certified testing methodologies as required by Arizona state law. Testing results were obtained according to Smithers' quality assurance plan and requirements found in R9-17-404.03 and R9-17-404.04. This COA is governed by the terms and conditions listed on: <https://www.smithers.com/arizona-terms-conditions>



Jeeter
2626 South Roosevelt Street
Tempe, AZ 85282
License #: 00000066DCBO00410690
Sample ID: 2508SMAZ1450.4260
Batch #: DFAZ-WATZKZ-082725



SMITHERS

CERTIFICATE OF ANALYSIS

License #: 00000020LCVT89602592

Certificate: 15991

Notes: Method of Extraction: Butane & Pressed



Ahmed Munshi

Technical Laboratory Director

Smithers CTS Arizona LLC
734 W Highland Avenue, 2nd Floor
Phoenix, AZ 85013
(602) 806-6930



Accreditation #: 103104

The product associated with this COA has been tested by Smithers CTS Arizona LLC, using validated state certified testing methodologies as required by Arizona state law. Testing results were obtained according to Smithers' quality assurance plan and requirements found in R9-17-404.03 and R9-17-404.04. This COA is governed by the terms and conditions listed on: <https://www.smithers.com/arizona-terms-conditions>